

The right way to measure.

Important tips on the operation and maintenance of the VITO® oiltester

Preparation







Do not touchRemove friedthe hot probe.goods before tDanger of burning!measurement.

Remove fried Wait until th

Remove fried Wait until there goods before the are no more bubbles rising (approx. 5 min.).



Maintain a minimum distance of 1 cm from metallic parts.



Measure in hot oil min. 40 °C, max. 200 °C.

Measurement



Switching the instrument on [O].



Observe min/max marks!



Immerse the sensor in hot oil. 1. Move sensor in oil.

Auto Hold activated:
Auto Hold in the display indicates the end of the measurement. TPM value and display colour flash until the final value is reached.

- Final value is fixed.
- 2. To return to measurement mode: [Hold] press briefly (< 1 s).



Record the determined measurement values.

Care & maintenance Cleaning





Clean instrument with mild detergent. Rinse sensor under running water and dry carefully with a soft paper towel.

Accuracy test

Your specific starting value:

%TPM

For commissioning your new instrument, measure in unused cooking oil, and note this TPM reference value (mean value from several measurements). This value serves as a comparative value for subsequent instrument testing.